



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/021,809	10/30/2001	Robert E. Simonson	N908	2099
21125	7590	04/21/2005	EXAMINER	
NUTTER MCCLENNEN & FISH LLP WORLD TRADE CENTER WEST 155 SEAPORT BOULEVARD BOSTON, MA 02210-2604			HO, UYEN T	
			ART UNIT	PAPER NUMBER
			3731	

DATE MAILED: 04/21/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No.	Applicant(s)	
	10/021,809	SIMONSON, ROBERT E.	
	Examiner	Art Unit	
	(Jackie) Tan-Uyen T. Ho	3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 02 March 2005.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 21-24 and 30 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 21-24, 30 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 3/2/05 has been entered.

Claim Rejections - 35 USC § 112

2. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

3. Claim 24 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. An implant, spinal rod is not positively claimed and it is not part of a cannula. The preamble recites a cannula but not an apparatus including a cannula and a spinal rod. It is not clear whether the applicant intends to claim an apparatus or just a cannula.

Claim Rejections - 35 USC § 102

4. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Art Unit: 3731

5. Claims 21-24 and 30 are rejected under 35 U.S.C. 102(b) as being anticipated by Ogawa et al. (5,976,146).

Ogawa et al. disclose a dilator retractor/cannula (Fig. 56D, 97A) as claimed. Wherein the retractor/cannula includes a main elongated body (56C, 402), a distal end portion (130, 403) being contoured to match the bone structure of a patient, the distal end being sloped including a short side and a long side and two opposed cut-out (arch shape in 132 portion, 414b in fig. 97A) between the short side and the long side of the distal end and defining a lateral passageway oriented at an angle to the working channel.

The distal end being shaped to define a lateral passageway oriented at angle to the working channel, the passageway is oriented generally perpendicular to the working channel and sized to pass a spinal rod (401, 402) there through. The spinal rod (401) is adapted to be inserted to the working channel from through the cut-out portion or the curve portion at the distal end.

The introductory statement of intended use and all other functional statements have been carefully considered but are deemed not to impose any structural limitations on the claims distinguishable over the Ogawa et al. 's device which is capable of being used as claimed if one desires to do so.

Claim Rejections - 35 USC § 103

6. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

Art Unit: 3731

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.


7. Claims 21-24 and 30 are rejected under 35 U.S.C. 103(a) as being unpatentable over Ogawa et al. '146. Ogawa et al. disclose cannula (Figs. 75B, 75C and 76B) having opposed cut-out formed in the distal end and defining a lateral passageway oriented at an angle to a working channel having slope or tunnel distal end and having tunnel funnel shaped (12, fig. 63A and 405, fig. 100A and 504, figs. 121, 132). Although, Ogawa et al. does not disclose in the embodiment disclosed in figures 75B, C and 76B a distal end being sloped such that the cannula has a short side and a long side, Ogawa et al. disclose through out the specification dilator/cannula with distal end being sloped to match the contour of a surgical site. Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to employ a sloped distal end into the dilator/cannula (211) of Ogawa et al. in order to provide a distal end having a contour that fit the surgical site.

8. Any inquiry concerning this communication or earlier communications from the examiner should be directed to (Jackie) Tan-Uyen T. Ho whose telephone number is 571-272-4696. The examiner can normally be reached on MULTIFLEX Mon. to Sat..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, ANHTUAN NGUYEN can be reached on 571-272-4963. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3731

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



(Jackie) Tan-Uyen T. Ho
Patent Examiner
Art Unit 3731

April 15, 2005